ICMR-National Institute of Cholera & Enteric Diseases P-33, C.I.T. Road, Scheme –XM, Beliaghata Kolkata – 700 010

AMENDMENT

Date: 10.11.2017

Last date of submission of Bid: 20.11.2017 (upto 1:30 PM). Last date of opening of Bid: 20.11.2017 (3:00 PM).

D-2/Equipment 2017-18 (32)

NAME : ULTRACENTRIFUGE

QUANTITY: 1 (ONE)

Amended Specifications

1. Floor standing

2. Fully microprocessor controlled

3. Maximum Speed (rpm): 100,000 or more

4. Speed Control accuracy: ± 10 (max) rpm of set speed (above 1,000 rpm)

5. Set Temperature: $+2^{\circ}$ C to $+40^{\circ}$ C

6. Temperature Control: $\pm 0.5^{\circ}$ C

7. Ambient Operating Range: 10° C to 35° C

8. Accel/Decel Profiles: 10/11

9. Timer: up to 99 hrs or more; HOLD

10. Suitable vacuum system for high-speed operation

11. User-Defined Programs: Yes12. Rotor Tracking: Yes

13. Built in imbalance protection: Yes14. Built-in LCD touch-screen display: Yes

15. Step-by-step zonal/CF operation screens: Yes

16. Drive Cooling: Air-cooled

17. Refrigeration System: Thermoelectric module; no CFC, HCFC, ODC etc.

18. Sound Level: <55 dBA

19. Power Requirement: 200-240VAC, 30 A, 50 Hz
20. Warranty: 5 year comprehensive
21. Servo Stabilizer: 5 KVA minimum

- 22. Titanium rotors (i) fixed angle 6 to 8 slots \times 30 ml for 50,000 to 80,000 rpm;
 - (ii) fixed angle rotor 8 to 10 slots \times 4 ml for \sim 80,000 rpm;
 - (iii) swing out rotor of 4 to 6 slots \times ~4 ml for 50,000 to 60,000 rpm
- 23. Optional rotor to be quoted: fixed angel for 1.5 ml conical tubes for virus work
- 24. All accessory items including tubes for each rotor should be quoted as optional items
- 25. Documents of International Certification to be provided
- 26. The vendor is expected to have full-fledged service set-up in Kolkata.
- 27. List of installations of the quoted model within last three years should be provided.
- 28. Performance Certificates from at least three users of government / government-aided research / academic institutes along with the installation certificates of the quoted model should be provided.
- 29. Compliance certificates should be included.

** Other terms & conditions and specifications will remain unchanged.

Administrative Officer for Director